1114

Select Scales Ltd

Sandy Lane Industrial Esate
Worksop Notts S80 1TN
Tel: 01909 722960 Fax: 01909 724057

CALIBRATION CERTIFICATE

CUSTOMER
SO Recycle Ltd
Unit 13/14 Romsey Ind. Estate
Greatbridge Road
Romsey
SO51 0HR

MACHINE DETAILS
Make: Dini Argeo
Model: 3590ET
Capacity: 50000kg
Division: 20kg
Serial no. 63202654
Location:

Certificate No. SORE1002

Test Date: 21st July 2020
Next Test Due: July 2021

The following certification tests have been carried out using calibrated weights traceable to the National Physical Laboratory Standards of Mass, uncertainties quoted provide a level of confidence of approximately 95%.
The certificate provided, is in accordance with the UK Weighing Federation Code of Practice.

Linearity and Hysteresis

Load	As Found Reading	As Found Deviation	Definitive Reading	Definitive Deviation
0T	0.00T	Nil		
2T	2.00T	Nil		
4T	4.00T	Nil		
8T	8.00T	Nil		
12T	12.00T	Nil		
16T	16.00T	Nil		
20T	20.00T	Nil		
38.36T	38.34T	-1		
43.40T	43.38T	-1		

Repeatability (x3) tested at:

Test Point	As Found Reading	As Found Deviation	Definitive Reading	Definitive Deviation
21.36T	21.36T	Nil		
21.36T	21.36T	Nil		
21.36T	21.36T	Nil		

Eccentric Loading

2 4 6
1 3 5

As Found Test

	Test Point	Reading	Deviation
Pos. 1	7.06T	7.06T	Nil
Pos. 2	7.06T	7.06T	Nil
Pos. 3	7.06T	7.06T	Nil
Pos. 4	7.06T	7.06T	Nil
Pos. 5	7.06T	7.04T	-1
Pos. 6	7.06T	7.06T	Nil
Pos. 7			
Pos. 8			

Definitive Test

	Test Point	Reading	Deviation
Pos. 1			
Pos. 2			
Pos. 3			
Pos. 4			
Pos. 5			
Pos. 6			
Pos. 7			
Pos. 8			

Test Weight Information

Cert no.	Weights
TR 10745	LWS 1-40

Remarks (To include weighing range if less than the maximum weighing capacity)

Fork Truck = 5.06T

Tolerances (MPE) Class III

0-500divs = ± 1 div
500-2000divs = ± 2 divs
2000-3000divs = ± 3 divs

Engineers ..... Justin Whitfield

Engineers Name (Print)..... Justin Whitfield

(F).29/1.5

This certificate is provided for internal purposes only